



UK NEQAS for Rubella IgM serology
Distribution : 1450
Dispatch date : 14-May-2001

Laboratory : 9999
Page 1 of 4

Intended Result	Your Report	Your Score
Specimen 5951 Rubella IgM Negative	Rubella IgM Negative	2
Specimen 5952 Rubella IgM positive	Rubella IgM positive	Not scored
Specimen 5953 Rubella IgM Negative	Rubella IgM Negative	2
Specimen 5954 Rubella IgM positive	Rubella IgM positive	2

Cumulative score information

Total number of specimens sent to you for **UK NEQAS for Rubella IgM serology** over the last 2 distributions is 8
For these distributions specimen numbers 5795 5796 5797 5798 5951 5953 5954 have been analysed and scored.

Number of reports analysed **7**
 Number of specimens reported as not examined (not scored) 0
 Number of specimens received too late for analysis (not scored) 0
 Number of specimens for which no report was received (scored as 0) 0
 Your cumulative score for these specimens was 14 out of a possible total of 14

The mean score calculated from the reports returned by UK laboratories was **13.77** (with a standard error of 1.14).
 Your performance rating for UK NEQAS for Rubella IgM serology (i.e. the number of standard errors by which your cumulative score lies above or below the mean) compared to UK laboratories is **0.20**

COMMENTS

● Comentários

Rácio do seu desempenho ●

Pontuação média e erro padrão ●

Informação sobre a sua pontuação acumulada ●

Nº de amostras pontuadas, incluindo amostras de relatórios não enviados ●

Resultados-alvo ●

Os seus resultados ●

A sua pontuação ●

O seu nº de participante ●